





# LiNA Xcise™

## Laparoscopic Morcellator





**Caution:** Federal Law (USA) restricts this device to sale by or on the order of a trained physician.

**Intended Use:**

The LiNA Xcise, Laparoscopic Morcellator is intended for gynaecologic use by trained professionals in hospital and surgical environments.

**Indications:**

Indicated for cutting, coring and extracting tissue in operative laparoscopy, including gynaecologic procedures such as hysterectomy and myomectomy.

**Contraindications:**

Contraindications for use on vascularised tissue (ovaries, fallopian tubes, myomas and other structures); must be devascularised before morcellation.

Laparoscopic power morcellators are contraindicated in gynecologic surgery in which the tissue to be morcellated is known or suspected to contain malignancy.

Laparoscopic power morcellators are contraindicated for removal of uterine tissue containing suspected fibroids in patients who are:

- peri- or post-menopausal, or
- candidates for en bloc tissue removal, for example, through the vagina or via a mini-laparotomy incision.

**Warnings:**

**Uterine tissue may contain unsuspected cancer. The use of laparoscopic power morcellators during fibroid surgery may spread cancer, and decrease the long-term survival of patients. This information should be shared with patients when considering surgery with the use of these devices.**

- The LiNA Xcise is provided STERILE. Carefully inspect the packaging for any damage prior to use. Do NOT attempt to use the device if sterile barrier is damaged. Do NOT use past expiration date.
- For single use only. Do NOT reuse, reprocess or re-sterilize the LiNA Xcise. Any reprocessing may impede the functions of this device. Reusing single use devices may also increase the risk of cross contamination. Attempts to clean the device results in risk of device malfunction and/or erroneous pathology specimen collection due to residual tissue in the LiNA Xcise causing significant gas leakage through the morcellator.
- Exercise caution while manipulating the LiNA Xcise. Do NOT place the cutting tip nearby or in contact with tissue which is not intended to be morcellated.
- Be aware that the cutting tip of the LiNA Xcise is NOT in contact with other instruments.
- Do NOT activate the LiNA Xcise if it is not possible to visualize the cutting tip.
- Do NOT attempt to sharpen or modify the cutting tube. Modified or distorted cutting tube can result in patient, surgeon or equipment damage.
- After use, dispose of product and packaging in accordance with hospital, administrative and/or local government policy. Be aware of sharp edges.

**Precautions:**

Use of the LiNA Xcise requires adequate training and experience in performing laparoscopic myomectomy and hysterectomy.

Be careful when inserting or removing the device. Make sure that the cutting blade is retracted by putting the trocar in the "Safety Guard" position during insertion and removal and whenever the cutting blade is not in active use. Insertion and removal of the LiNA Xcise should always be performed under direct visual control. Keep the rotating blade visible during the entire morcellation procedure.

Failure to carefully follow all applicable instructions may result in significant injury to the patient, physician or attendants and may have an adverse effect on the outcome of procedures performed.

**Product description:**

The LiNA Xcise contains a rotating cutting tube with built-in trocar function that also serves to protect the sharp end of the cutting tube. A grasper or a tenaculum forceps must be used to pull the stripes of tissue through the lumen of the cutting tube. The LiNA Xcise cutting function is controlled by the activation button on the hand piece.

**Instructions for use:**

The surgeon should read the Instructions for use carefully before using this device.

1. Carefully inspect the packaging for any damage prior to use. Do NOT attempt to use the device if sterile barrier is damaged.
2. Prior to using the trocar function of the LiNA Xcise, insert the obturator fully into the device. Be sure that the trocar is placed in the "Safety Guard" position. If not, place the trocar in the "Safety Guard" position by holding in the bracket and then turn the trocar (see illustration in picture 2).
3. The LiNA Xcise with obturator should be placed into the abdomen using standard technique for laparoscopic trocar placement. It is recommended to insert the LiNA Xcise with obturator through a 12-14mm incision under direct visualization.
4. In order to remove tissue use a 10-12 mm forceps or similar instrument inserted through the lumen of the LiNA Xcise and into the abdomen. To prevent injury to the abdominal wall, the tissue to be morcellated should be completely exposed before attempting to extract it through the morcellator.
5. It is recommended to use a second pair of grasping forceps to hold the tissue in place and reduce tissue movement during morcellation.
6. Place the trocar in the required position by turning the trocar into cutting position: "Cut 1" for peeling function or "Cut 2" for coring function.
7. Adjust the coreguard if needed.
8. To activate the cutting blade and begin morcellating, press the activation button on the hand piece while pulling pieces of tissue through the cutting tube.
9. Release the activation button as soon as the stripe of tissue is extracted from the LiNA Xcise.
10. For use with a 5 mm instrument: Mount the reducer cap onto the back of the morcellator with a pressing and turning motion.
11. After surgery, remove the LiNA Xcise from the abdominal cavity. For proper disposal, turn the trocar into the "Safety Guard" position, remove the reducer cap and insert the obturator. The morcellator may now be safely disposed in accordance with local governing ordinances and recycling plans.

Country of origin: Poland

US